

Posi3 USB Calibration Report

Serial Number U08195

2022-10-14

Zero Stability		<i>Before</i>	<i>After</i>	
	Last Cal	Calibration	Calibration	Tolerance
FACEPIECE	2021-9-20	0,00	0,00	± 0,1 "H2O
ACTIVATION	2021-9-20	0,00	0,00	± 0,3 "H2O
MEDIUM	2021-9-20	0,0	0,0	± 1 PSI
HIGH 2	2021-9-20	0	0	± 10 PSI

Instrument was in tolerance
before calibration

☒ YES ☐ NO

☐ Could not be determined

Facepiece Pressure		<i>Before</i>	<i>After</i>	
Applied	Calibration	Calibration	Calibration	Tolerance
0 "H2O	-0,01	0,00	0,00	± 0,1 "H2O
3 "H2O	2,98	2,99	2,99	± 0,1 "H2O
5 "H2O	4,95	4,97	4,97	± 0,1 "H2O

The signing of this document indicates that this instrument's calibration has been verified and meets the specifications listed on this document. The results from these readings have also been documented on this sheet.

Activation Pressure		<i>Before</i>	<i>Working</i>	<i>After</i>	<i>Calibration</i>
Applied	Calibration	Tolerance	Calibration	Calibration	Tolerance
-20 "H2O	-20,13	± 0,5 "H2O	-20,27	-20,27	± 0,3 "H2O
-10 "H2O	-10,01	± 0,5 "H2O	-10,10	-10,10	± 0,3 "H2O
0 "H2O	0,01	± 0,5 "H2O	0,00	0,00	± 0,3 "H2O

Technician's Signature

Lung		<i>Before</i>	<i>After</i>	
	Last Cal	Calibration	Calibration	Tolerance
STD	2021-9-20	1,66	1,67	± 0,05 L
CUST	2021-9-20	2,48	2,50	± 0,05 L
MAX	2021-9-20	3,35	3,39	± 0,05 L

Medium Pressure		<i>Actual</i>		<i>Actual</i>		
Applied	<i>Applied</i>	<i>Before</i>	Working	<i>Applied</i>	<i>After</i>	Calibration
(± 25 PSI)	<i>Pressure</i>	<i>Calibration</i>	Tolerance	<i>Pressure</i>	<i>Calibration</i>	Tolerance
* 100	100,0 PSI	100,8	± 3 PSI	100,0	100,7	± 1 PSI
200	200,0 PSI	202,0	± 3 PSI	200,0	200,1	± 1 PSI

2nd High Pressure		Actual		Actual		
Applied	Applied	Before	Working	Applied	After	Calibration
(± 100 PSI)	Pressure	Calibration	Tolerance	Pressure	Calibration	Tolerance
1000	1000 PSI	993	± 12,5 PSI	1000	997	± 10 PSI
2000	2000 PSI	1991	± 12,5 PSI	2000	1998	± 10 PSI
3000	3000 PSI	2992	± 12,5 PSI	3000	3001	± 10 PSI
4000	4000 PSI	3990	± 12,5 PSI	4000	4002	± 10 PSI
4600	4600 PSI	4595	± 12,5 PSI	4600	4607	± 10 PSI

Calibrated By:
Lucas Tripodi
Pichincha 1580

Test Credits Before Cal 50000
Test Credits After Cal 50000

Customer:
WELL DONE SERVICES SRL
Pte Hipolito Yrigoyen 416,5to A, Neuquen
alfredo.moretti@welldonesrl.com.ar

